

Multi-wavelength Calibration Procedure for the Pierre Auger Observatory Fluorescence Detectors

A.C. Rovero^{*,a,1}, P. Bauleo^b, J.T. Brack^b, J.L. Harton^b, and
R. Knapik^b, for the Pierre Auger Collaboration

^a*Instituto de Astronomía y Física del Espacio (CONICET), CC 67 Suc 28
(C1428ZAA) Buenos Aires, Argentina*

^b*Colorado State University, Department of Physics, Fort Collins CO 80523,
USA*

Abstract

We present a method to measure the relative spectral response of the Pierre Auger Observatory Fluorescence Detector. The calibration was done at wavelengths of 320, 337, 355, 380 and 405 nm using an end-to-end technique in which the response of all detector components are combined in a single measurement. A xenon flasher and notch-filters were used as the light source for the calibration device. The overall uncertainty is 5%.

Key words: Auger Observatory, Extensive Air Shower, Fluorescence detectors, calibration, cosmic ray

PACS: 95.55.Cs, 96.50.sd

*Corresponding author. e-mail: rovero@iafe.uba.ar

¹Member of Carrera del Investigador, CONICET

1. Introduction

The Pierre Auger Observatory has been designed to measure Extensive Air Showers (EAS) initiated by cosmic rays with energies above 10^{18} eV. The Observatory calls for the construction of two large detectors, one in the southern hemisphere and one in the northern hemisphere, each covering an area of at least 3000 km^2 [1]. The Southern Observatory original baseline design in Malargüe, Argentina, is completed and consists of two detectors, the Surface Detector (SD) and the Fluorescence Detector (FD). The SD is composed of 1600 water Cherenkov detectors located on a triangular array of 1.5 km spacing to measure the EAS secondary particles reaching ground level. In addition, the UV-nitrogen fluorescence light produced in air is registered by the FD during dark, clear nights. The FD consists of 24 telescopes distributed in four buildings, or FD stations, overlooking the SD array.

The energy calibration of data taken at the Pierre Auger Observatory relies on the calibration of the FD [2] [3]. A detailed description of the fluorescence detector can be found elsewhere [4]. The Auger FD telescopes use Schmidt optics. The aperture is defined by a 2.2 m optical diaphragm. A UV filter covers the aperture and reduces background light by cutting out all light not in the main part of the nitrogen fluorescence spectrum (~ 300 - 400 nm). It also provides ambient isolation, which allows for temperature controlled operation of the telescope and prevents dust from entering the optical system. Spherical aberrations are reduced by a Schmidt corrector annulus covering only the outer portion of the aperture. Light is concentrated by a $3.5 \text{ m} \times 3.5 \text{ m}$ tessellated spherical mirror into an array of 440 hexagonal photomultipliers (PMTs), referred to as “pixels”, with a field of view of 1.5° each. At the focal plane, light concentrators approximating hexagonal Winston cones reduce dead spaces between PMTs. The pixel array is referred to as a “camera”. In Fig. 1 we show the main components of the FD telescope.

To calibrate the FD three different procedures are performed [5]: the relative, the absolute and the multi-wavelength calibrations. Relative calibration is performed at least at the beginning and at the end of every observing night. It is based on uncalibrated but stable light sources that illuminate the camera from three positions upstream in the optical system, tracking the nightly response variations of the whole system [6]. The absolute calibration is made by an end-to-end technique, using a calibrated portable light source in front of the telescope aperture, which calibrates the combined effect of each component in a single measurement at a single wavelength, 375 nm .




Figure 1: Sketch of the fluorescence detector showing its main components, together with the calibration light source (drum) in calibration position. PMT camera supports and drum supports are not shown for clarity.

38 The light source has been designed to uniformly illuminate all 440 pixels in
 39 a single camera simultaneously and is referred to as the “drum” because of
 40 its appearance. It is a cylinder of 2.5 m diameter and 1.4 m deep, with one
 41 Teflon[®] face, and internally laminated with Tyvek[®] (see Fig. 2). When
 42 used for absolute calibration of FD telescopes, a UV LED is placed inside
 43 a small Teflon diffuser inside the drum, and surrounded by other diffusive
 44 pieces in such a way that the face is uniformly illuminated. The procedure
 45 to calibrate the drum at the laboratory has been outlined elsewhere for the
 46 prototype [7] and for the current version of the drum [8]. Absolute calibra-
 47 tion of FD telescopes is performed typically twice a year to follow long term
 48 variations of the system response. Finally, a “multi-wavelength calibration”
 49 procedure determines the spectral response of the system as a function of
 50 photon wavelength. This is a relative measurement, normalised to the abso-
 51 lute calibration at 375 nm. The multi-wavelength calibration is needed not
 52 only for correct event reconstruction but also to correlate with the results
 53 of alternative absolute calibrations performed at different wavelengths using
 54 lasers. Changes in the spectral response of the FD are not expected to oc-
 55 cur in the short term, thus the frequency for evaluating this dependence is
 56 planned to be less than once per year.

57 In this work, we describe the procedure for multi-wavelength calibration of

58 fluorescence telescopes using an end-to-end technique similar to that used for
59 the absolute calibration. We describe the initial spectral dependence function
60 used by the Auger Observatory in section 2. The new light source used in
61 the procedure is described in section 3 and its characterisation in section 4.
62 The FD-telescope wavelength response and a discussion of uncertainties are
63 presented in section 5.

64 2. Piecewise spectral response of the fluorescence detector

65 The spectral response of the FD originally used by the Auger Observatory
66 was assembled from the efficiencies of the individual telescope components.
67 The individual efficiencies were obtained from statements by the component
68 manufacturers or, in some cases, as measured by members of the Pierre
69 Auger Collaboration [9]. The elements considered for the spectral response
70 were the UV filter and corrector ring transmission, the mirror reflectivity, and
71 the PMT quantum efficiency. The overall wavelength response is dominated
72 by filter and PMT effects. We call this piece-wise curve $PW(\lambda)$ and show it
73 in section 5 to compare it with results in Fig. 7.

74 Regardless of the accuracy of the measurements on each telescope com-
75 ponent, the fact that the response of the system was not measured as a whole
76 indeed introduces uncertainties. We can calculate, for example, that the cor-
77 rector ring transmission does not affect all the incoming photons but only
78 approximately half of them, when camera shadowing effects are included.
79 Also, the reflectivity of the light concentrators in the camera are not taken
80 into account at all, even though a third of the light getting to the pixel is
81 reflected by them. There are other less important considerations such as
82 the fact that mirrors for the FD telescopes are made using two different
83 techniques by two different manufacturers [10]. One of them is a machined
84 aluminium alloy with a protecting layer of Al_2O_3 at its reflecting surface. In
85 the second technique polished glass is aluminized and the reflecting surface
86 is covered by a layer of SiO_2 . Both types of mirrors have similar reflectivities
87 as measured at 370 nm [10]. However, considering that they use different di-
88 electric coatings on different materials, the reflectivity at other wavelengths
89 may vary.

90 To assure that the Pierre Auger Observatory is using the right spectral
91 response of its FD telescopes, the decision was made to adapt the end-to-end
92 procedure used for absolute calibration to directly measure this function.

93 3. Multi-wavelength light source

94 To enable multiple wavelength measurements, the LED used for absolute
95 calibration was removed and a light pipe was installed between the Teflon
96 diffuser and the back of the drum, where new light sources could be mounted
97 (see Fig. 2). A xenon flasher is mounted at the end of the pipe at the
98 back of the drum. The xenon flasher² provides 0.4 mJ optical output per
99 pulse covering a broad UV spectrum, in a time period of a few hundred
100 nanoseconds. To select a desired wavelength, notch-filters³ are mounted in
101 a filter wheel attached to the end of the pipe. A focusing lens at the filter
102 wheel output maximises the intensity through the filter wheel and into the
103 light pipe. Notch-filters were chosen at five wavelengths inside the range
104 of the FD UV filter located at the telescope aperture. According to the
105 manufacturer, the filter transmissions are centred at 320, 337, 355, 380 and
106 405 nm, with a FWHM ≈ 15 nm.




Figure 2: Drum section showing the main components of the light source. The diffusively reflecting surfaces a) and b) are Teflon, while c) is Tyvek

²RSL-2100 series xenon flasher - Perkin Elmer, www.perkinelmer.com

³Standard UV bandpass filters, Andover Corporation, www.andcorp.com

107 4. Characterization Measurements

108 The drum relative intensity is measured at each wavelength in a dedicated
109 calibration laboratory at the Observatory. An auxiliary PMT, the “lab-PMT”,
110 is used to measure the light intensity of the drum for a given notch-filter.
111 Quantum efficiency (QE) of the lab-PMT and FD spectral response have
112 significant variations within the range of wavelengths where the notch-filters
113 transmit. To understand the corrections to be applied due to these variations
114 we have performed measurements described in the following sections.

115 4.1. Notch-filter transmission scan

116 Notch-filters listed in section 3 were selected at 5 wavelengths either near
117 the main emission bands of nitrogen [11] used for EAS fluorescence detection
118 (320, 355 and 405 nm), or to match the laser wavelengths (337 and 355 nm)
119 used in previous absolute calibration cross checks using roving lasers [5], or
120 near the 375 nm LED single-wavelength absolute calibration. Precise mea-
121 surements of the transmission characteristics of each notch-filter are needed.
122 The filters were scanned using a monochromator with a broadband deuterium
123 light source. A photo-diode of known wavelength dependence [12] was used
124 at the monochromator output to detect the transmitted light as a function of
125 wavelength in 2 nm steps. The results are shown in Fig. 3 where the spectral
126 shape of the deuterium light source and the response of the photo-diode have
127 been deconvolved. We assign labels of $f_i(\lambda)$ to the curves in the figure and λ_i
128 to the central wavelength of each of them, where $i = 1, 5$ indicates one of the
129 five notch-filters (from 320 nm to 405 nm). A scan from the manufacturer
130 was available for one of the filters (337 nm), and it was found to be in good
131 agreement with our scan. Some asymmetries were found in the transmission
132 curves (320 and 337 nm) that make significant differences when applying
133 filter corrections.

134 4.2. Quantum efficiency of the lab-PMT

135 We measured the relative quantum efficiency of the lab-PMT in 2 nm
136 steps using the deuterium light source and the monochromator. We directed
137 the monochromator output through a thin Teflon diffusor and into a dark
138 box containing the lab-PMT and a NIST-calibrated photo-diode⁴ [12]. Using
139 the small photo-diode we verified that the beam was uniform laterally at a

⁴UV100 photodiode, UDT Sensors, Inc, Hawthorne, CA USA




Figure 3: Relative transmission of the notch-filters used for multi-wavelength calibration. The nominal wavelength is indicated for each filter

level of 0.5% over an area larger than the PMT photocathode. The first step in the QE measurement was to scan the monochromator and measure the relative output intensity in photons at each wavelength, using the photodiode and its known calibration. Then, using an iris to limit the intensity and prevent PMT saturation, we rescanned the source and measured the PMT current. The PMT relative QE as a function of wavelength, $QE(\lambda)$, is the ratio of these two scan results at each wavelength.

The measured QE, shown in Fig. 4, is in agreement with the average photocathode QE provided by the manufacturer⁵, within the measurement uncertainties of 2.5%, which are discussed in section 5.2.

4.3. Drum Intensity at five wavelengths

With the notch-filter wheel mounted on the drum, the relative intensity of the drum surface for each wheel position depends on the xenon source intensity at the transmitted wavelengths, the notch-filter transmission and losses in the light pipe. No wavelength shifting of photons of Teflon and Tyvek

⁵78 mm, 9265B PMT, made by Electron Tubes, www.electrontubes.com




Figure 4: The measured quantum efficiency of the lab-PMT used in this work, and the typical QE from the manufacturer’s specification sheet

155 materials has been observed in our independent laboratory measurements.
156 Ideally, to measure the drum intensity for each wavelength, one would use
157 the drum surface as input to the monochromator and scan the full spectrum
158 for each notch-filter. In practice this is precluded by the low drum intensity.
159 Instead, we make a single measurement of the integrated drum intensity for
160 each notch-filter using the lab-PMT. For each notch-filter we find a value
161 for C_i , the centroid of the histogram of the PMT response to 1000 xenon
162 flasher pulses. These centroids are proportional to the drum intensity for
163 each notch-filter once corrections have been made for variations in lab-PMT
164 QE. Ignoring common constants, $\int \Phi_i(\lambda) QE(\lambda) d\lambda = C_i$, where $\Phi_i(\lambda)$ is
165 the brightness of the drum surface for the notch-filter i as a function of
166 wavelength. Then, since the distribution of drum photons is the convolution
167 of the known xenon flasher spectrum, $Xe(\lambda)$, and the corresponding notch-
168 filter, we use $\Phi_i(\lambda) = k_i f_i(\lambda) Xe(\lambda)$ and adjust k_i to match the integral
169 above. With this last process all five values of drum brightness and their
170 wavelength distributions are known.

171 The quantity $\int \Phi_i(\lambda) d\lambda = \Phi_i$ is proportional to the real total photon flux
172 being emitted by the drum surface. All the wavelength independent prop-
173 erties (PMT gain, electronic conversion, etc.) are left out because they will

cancel in the end when the relative values are computed. We also note that Φ_i is not significantly different from the value it would have if the function $QE(\lambda)$ was totally flat within the notch-filter range. Then, in practice,

$$\int \Phi_i(\lambda) d\lambda = \Phi_i \simeq C_i/QE(\lambda_i) \quad (1)$$

5. Fluorescence Detector response to drum

For testing the procedure we use results of measurements made at one FD telescope. In August 2006 we mounted the drum with xenon flasher and filter wheel at the aperture of telescope 4 at the Los Leones FD-building. A series of 400 xenon flashes illuminated the camera, and we found the average integrated pulse for each pixel. In Fig. 5 we show the response of one FD pixel and the distribution of pulse integrals for the same pixel. The pulse shape from the drum with the xenon source is irregular and varies from pulse to pulse, but the total output energy is consistent as indicated by the $<10\%$ RMS of the integral distributions. We obtained the average charge from the distributions of those 400 pulse integrals, $I_{i,j}$, for each notch-filter i and each pixel j . Typical statistical uncertainty for these values is $<0.5\%$.

The relative wavelength-dependent FD response for pixel j is then the ratio of the integrated ADC response of the FD to the relative number of photons at the aperture. We call this value $R_{i,j}$ so that $R_{i,j} = I_{i,j}/\Phi_i$. A considerable dispersion of $R_{i,j}$ values is expected as pixels have different gains and amplifications. However, we are only interested in relative wavelength responses so, all values are normalized to the response to the 380 nm notch-filter, or $i = 4$. Thus $R_{i,j}^{rel} = R_{i,j}/R_{4,j}$. In Fig. 6 the distribution of these relative responses for pixels in the measured telescope is shown for each notch-filter, except the one at 380 nm as this would be the identity. The histograms shown for each notch-filter in Fig. 6 have relatively low dispersions, ranging from 1.1% to 2.2% RMS. We consider that the relative response for each PMT in the camera is well represented by the average of those distributions, so only one value for each notch-filter is taken. We call these values R_i^{rel} , where i identifies the filter.

In Table 1 we show the R_i^{rel} values obtained for the telescope as evaluated from the distributions on Fig. 6. Uncertainties to these values are discussed in next sections. The resulting wavelength dependent efficiency is shown in Fig. 7 where the fitting was done adjusting the piecewise function to the




Figure 5: FD single pixel response to drum pulses. The signal of pixel number 220 of telescope 4 with drum at 355 nm is shown. Left: Distribution of pulse integrals related to the average integral. Right: Pulses as registered by the FD. Dots are the values for 400 individual pulses. The grey solid line is the average

207 five measured points R_i^{rel} . The piecewise response in its original form is also
 208 shown in the figure for comparison purposes. The fit has been corrected for
 209 notch-filter width effects, as described in next section.

210 5.1. Notch Filter Width Effects and fitting procedure

211 In section 4.3 we described how the drum centroids for five wavelengths
 212 were measured and corrected for the lab-PMT quantum efficiency to get the
 213 relative drum intensities. In that process the change in the QE was taken
 214 into account by considering the distribution of photons for each notch-filter.
 215 Then, in section 5, we described how the FD responses to those intensities
 216 were measured for one FD telescope. In this last process the notch-filter
 217 width effect was not taken into account. Because the overall FD response is
 218 not flat in the ~ 15 nm FWHM range of each filter, this uncorrected result is
 219 biased toward the region of the filter corresponding to higher FD response.

220 To correct for this effect we follow a similar procedure as in section 4.3.
 221 The process at the telescope is:

$$\frac{\int \Phi_i(\lambda) FD(\lambda) d\lambda}{\int \Phi_i(\lambda) d\lambda} = R_i^{rel} \quad (2)$$




Figure 6: Distribution of relative FD response for each notch-filter. The response relative to 380 nm of all pixels in one telescope is shown

222 where $FD(\lambda)$ is the FD relative response as a function of wavelength.
 223 Again, we are not including wavelength independent constants as this is a
 224 relative measurement procedure. Note that if the notch-filter transmission,
 225 $f_i(\lambda)$, were delta functions at λ_i , we would have $FD(\lambda_i) = R_i^{rel}$. To solve the
 226 integral of equation 2 and find the unknown function $FD(\lambda)$ we made the
 227 assumption that the FD response is a piecewise-like function, as described
 228 in section 2. This choice is required because the five measurements in the
 229 wavelength range of the FD response are not sufficient to fully characterise
 230 a generic function. Then, it is reasonable to adjust the original piecewise
 231 function, $PW(\lambda)$, to what was measured. This assumption also implies that
 232 we have taken the FD response as null beyond the wavelength of the FD

Wavelength	R_i^{rel}	
(nm)	Mean	Corrected
320	0.483	0.409
337	0.843	0.832
355	1.007	0.998
380	1.000	1.044
405	0.229	0.243

Table 1: Average relative FD response measured for each notch-filter. Values are taken from the distributions of Fig. 6. The column on the right have the values after the correction for notch-filter width effects as described in section 5.1

233 system. Particularly, we take $FD(280\text{ nm}) = FD(425\text{ nm}) = 0$, which have
234 been verified by measuring the UV-filter transmission in our laboratory.

235 The procedure of adjusting the fitting function to the measured points was
236 as follows. For each notch-filter we take $FD(\lambda) = h_i Fit(\lambda)$, where $Fit(\lambda)$
237 is the fitting function and h_i a parameter for filter i . As a first guess the
238 piecewise function is taken so, $Fit(\lambda) = PW(\lambda)$. Equation 2 is evaluated and
239 the parameter h_i found to match the value R_i^{rel} . Once all five h_i are found,
240 a new fit is done by interpolating the points $h_i Fit(\lambda_i)$ with a piecewise-like
241 function, where now i runs for seven points, including the null extremes.
242 Between these points, the curve is adjusted by a linear interpolation of the
243 adjustments at the surrounding points. Finally, this last fit is taken as a new
244 $Fit(\lambda)$ function and the process starts again until all five values h_i are the
245 identity.

246 The final result of the iteration procedure, $Fit(\lambda)$, is the $FD(\lambda)$ that
247 fulfils the integral in equation 2 for all five measured points. This curve is
248 shown in Fig. 7, normalised to the value at 380 nm. In the same figure we
249 also show the original piecewise function. A decrease in the spectral response
250 compared to the piecewise response is observed at shorter wavelengths, the
251 largest difference of $\approx -28\%$ comes at 320 nm. The corrected R_i^{rel} values are
252 shown in Table 1.

253 5.2. Uncertainties

254 Contributions to the uncertainty include those from measurements in the
255 laboratory of the drum intensity, and from measurements at the telescope
256 of the FD response. The overall result reported here is a relative measure-
257 ment, and consequently many factors cancel particularly systematics related




Figure 7: The measured Fluorescence Detector wavelength response compared with a curve generated in a piecewise fashion from manufacturer’s data for each FD component. The corrected measurements at the 5 wavelengths are shown; the solid curve is constrained to pass through those points

258 to laboratory setup.

259

260 The determination of drum intensity at each wavelength includes mea-
261 surements in the laboratory of the centroid of the lab-PMT response to pulsed
262 drum illumination, and the relative QE of the lab-PMT. The uncertainty in
263 the lab-PMT drum response centroid, C_i in equation 1, is estimated to be 1
264 channel in the ADC converter plus the statistical uncertainty on the mean
265 of the 1000 xenon pulse distribution. The 1 channel uncertainty is a system-
266 atic effect, related to repeatability, and has more relevance for wavelengths
267 where the drum brightness is low. The second column in Table 2 indicates
268 the uncertainties in C_i .

269 Uncertainties related to measurement of the relative QE of the lab-PMT
270 include those from lab-PMT response, photodiode current during monochro-
271 mator scans, the calibration of the photodiode at each wavelength, and sys-
272 tematics in the laboratory setup. The uncertainty in the monochromator

Wavelength (nm)	C_i (%)	Φ_i (%)	R_i^{rel} (%)		
			Statistical	Systematic	Total
320	0.6	2.6	2.2	4.1	4.7
337	1.0	2.7	1.5	4.2	4.5
355	2.0	3.2	1.1	4.5	4.6
380	2.0	3.2	n/a	0.0	0.0
405	0.3	2.5	1.3	4.1	4.3

Table 2: Sources of uncertainties and their values for this work (see section 5.2).

wavelength has been measured with a N2 laser light source to be less than ~ 0.25 nm, and no contribution to the overall uncertainty has been included for this effect.

For the PMT QE scans, currents from the photodiode and the lab-PMT were measured with electronics based on an integration chip with linearity of 0.005% [13]. Effects of connectors and cabling between the detectors and the integration chips are expected to dominate any nonlinearities. While these effects are expected to be small, we assign an overall uncertainty of 2% to current measurements, based on our experience measuring absolute currents using similar configurations. For this relative measurement, the experimental configuration remained unchanged during measurements at different wavelengths, allowing some readout systematics to cancel in the final ratios. The uncertainties in the absolute response of the photodiode, as calibrated by NIST, are documented as k=2 expanded uncertainties [12]. Assuming normal distributions for the contributing factors, these can be interpreted as 1-sigma uncertainties varying in the range from 0.55 to 0.95% for wavelengths between 300 and 400 nm, and 0.4% or less for those between 405 and 450 nm. Monochromator output beam uniformity has been measured in the PMT-photodiode region to be better than 0.5% across a 10 cm region perpendicular to the beam. No contribution to the overall uncertainty has been included for beam non-uniformity. The stability of the system, including the lab-PMT response, was tested by performing measurements multiple times to find that repeatability is well within 1%. As a result of these effects we consider an overall uncertainty of 2.5% for the relative QE of the lab-PMT. Combination of the uncertainties of C_i and QE give the uncertainty in the drum brightness, Φ_i , as listed in Table 2.

300 The camera-averaged response of pixels to each of the five wavelengths
 301 are described in section 5. To evaluate the uncertainties in these responses
 302 we consider the systematics coming from drum fluxes, Φ_i as discussed above,
 303 and the widths of the distributions of responses for the 440 pixels in the
 304 camera, as taken from the distributions of Figure 6 and shown in the fourth
 305 column of Table 2. For a given pixel j , $R_{ij}^{rel} = R_{ij}/R_{4j}$, four quantities have
 306 to be considered: two integrals, I_{ij} and I_{4j} , and two drum fluxes, Φ_i and
 307 Φ_4 , as defined in previous sections. Uncertainties in the integrals are purely
 308 statistical and have been evaluated to be $<0.5\%$. The combination of these
 309 four values gives the total uncertainty shown in the fifth column of Table 2.
 310 Values at different wavelengths are independent and are added in quadrature,
 311 except at 380 nm where systematic uncertainties are null by definition. The
 312 last column in Table 2 shows the combined uncertainty for the two sources.
 313

314 Finally, we consider uncertainties resulting from the fitting procedure
 315 incorporating the effects of the notch-filter widths, as described in section
 316 5.1. The shape of the piecewise calibration curve, used as the initial fitting
 317 function, is dominated near 300 nm by the fall-off of FD PMT quantum
 318 efficiency at shorter wavelengths, and by decreasing transmission of the UV
 319 filter through the FD aperture above 400 nm. We have tried several shapes
 320 for this initial function and found that any reasonable choice incorporating
 321 these two features leads to near-identical results. No related uncertainty is
 322 included.

323 When applying the notch-filter width correction we use the relative trans-
 324 mission of the 5 notch filters, measured as described in section 4.1. Associ-
 325 ated uncertainties are 0.5% at each wavelength in the scans. Good correlation
 326 with the manufacturer's scan available for one filter supports this value of
 327 the uncertainty. For an independent evaluation of the systematics related
 328 to filter shape, we took the wavelength with the largest filter-width effect,
 329 320 nm, and changed the filter shape to an extreme-case scenario of a step
 330 function. We included a maximum spread in intensity between two wave-
 331 lengths by assuming extremes of the photodiode uncertainties at the respec-
 332 tive wavelengths. We then repeated the notch filter correction calculation.
 333 The corrected value of R_1^{rel} given in Table 1 changed by -0.5% , supporting
 334 the assesment above.

335 Based on the discussions above, we assign an overall total uncertainty
 336 of 5% to the measurements reported here. The main contributions come
 337 from the uncertainties on the relative QE of the lab-PMT and the measured

338 relative drum fluxes at each wavelength.

339 6. Conclusions

340 The method for measuring the relative wavelength-dependence response
341 of the Pierre Auger fluorescence detector has been tested in one telescope.
342 Within uncertainties we can say that a multi-wavelength calibration for each
343 PMT in a given camera is not necessary as the dispersion around the average
344 is of the order of few percent.

345 This result indicates lower FD efficiency at shorter wavelengths when
346 compared to a curve constructed in a piecewise manner from manufacturer
347 efficiency specifications for the individual elements of the system. The piece-
348 wise curve was adjusted to the measured values and the result is currently
349 used in the Auger event reconstruction software.

350 References

- 351 [1] J. Abraham [Pierre Auger Collaboration], NIM A 523 (2004) 50-95.
- 352 [2] M. Roth [Pierre Auger Collaboration], Proceeding 30th ICRC, Merida
353 (2007).
- 354 [3] J. Abraham [Pierre Auger Collaboration], Physical Review Letters 101
355 (2008), 061101.
- 356 [4] V. Verzi [Pierre Auger Collaboration], Nucl. Phys. B (Proc. Suppl.) 165
357 (2007) 37.
- 358 [5] R. Knapik [Pierre Auger Collaboration], Proceeding 30th ICRC, Merida
359 (2007).
- 360 [6] C. Aramo [Pierre Auger Collaboration], Proceeding 29th ICRC, Pune
361 (2005), 8, 101.
- 362 [7] J. T. Brack, *et al.*, Astropart. Phys. 20 (2004) 653.
- 363 [8] P. Bauleo [Pierre Auger Collaboration], Proceeding 29th ICRC, Pune
364 (2005), 8, 5-8.

- 365 [9] J. A. J. Matthews [Pierre Auger Collaboration], Proceeding SPIE As-
 366 tronomical Telescopes and Instrumentation, Waikoloa , Hawaii (2003),
 367 4858, 121-130.
- 368 [10] G. Matthiae [Pierre Auger Collaboration], Proceeding 27th ICRC, Ham-
 369 burg (2001).
- 370 [11] M. Nagano, *et al.*, Astropart.Phys. 20 (2003) 293.
- 371 [12] T.C. Larson, S.S. Bruce, and A.C. Parr, “Spectroradiometric Detector
 372 Measurements”, National Institute of Standards and Technology, Cal-
 373 ibration Program, Gaithersburg, MD 20899-2330, Special Publication
 374 250-41, 1998.
- 375 [13] Burr-Brown Corporation, chip ivc102; <http://www.burr-brown.com/>